

ELECTRON PARAMAGNETIC RESONANCE (EPR) DOSIMETRY

Prepared by
The 4th years student
Of thr FEM, gr. 21051
Savin Andrey

ELECTRON PARAMAGNETIC RESONANCE (EPR) DOSIMETRY


Electron paramagnetic resonance (EPR) dosimetry is a physical method for the assessment of absorbed dose from ionising radiation. It is based on the measurement of stable radiation induced radicals in human calcified tissues (primarily in tooth enamel).




RESONANCE CONDITION

$$h\nu = g\mu_B B$$

- ν is resonance frequency
- h is Plank's constant
- g is the g-factor
- μ_B is the Bohr magneton
- B is the magnetic field induction



EPR SPECTROMETER



EPR spectrometer «Bruker»

CLASSES OF EPR SPECTROMETERS

- Large research spectrometers
- Middle-class spectrometers
- Small spectrometers



ADVANTAGES AND DISADVANTAGES OF THE EPR METHOD

+

- measure the volume of samples;
- dose reconstruction to the distinctive tissues;
- dose reconstruction after long periods of exposure;
- dose reconstruction for many years after the exposure.

-

- the difficulty in collecting material for analysis;
- reconstruction of the individual dose is complicated and labour-consuming.

THANKS FOR ATTANTION!!!